EQUIPMENT REQUIREMENTS

As per Revised BDS Course Regulations, 2007

Department: Oral Medicine and Radiology

	Specification	Qua	ntity	Available	Adequate /
Name		50 Intake	100 Intake	Quantity	Inadequate /
	(Space allocation for each dental chair 100 Sq.Ft. in all the Departments)				
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventry Suction, X-Ray viewer, 3 way syringe, instrument tray, Dental Operator's Stool with height adjustment		12	6	Adequate
with Cephalometric	Preferably digital	1	1	1	Adequate
ntra Oral X-ray Unit	70 KV, 8 mA, high frequency preferably Digital timer	2	3	2	Adequate
Pulp testers	Digital			2	Adequate
Automatic periapical X-ray Developer		1	1	2	Adequate
Automatic Panoramic with Cephalometric X- ray Developer		1	1	1	Adequate
X-ray viewer	For Panoramic and Cephalometric films	2	2	5	Adequate
Radiovisiography	Digital Intra X-ray system with one Sensor and Software	1	1	1	Adequate
General X-ray Unit		1	1	1	Adequate
Ortho Pantmograph	Preferably digital	1	1	1	Adequate
automatic Developers / Dark Room		1	2	1	Adequate
Lead aprons		2	2	4	Adequate
Lead Gloves		1	1	2	Adequate
X-ray Hangers		6	6	91	Adequate
X-ray Viewers		2	2	5	Adequate
Diagnostic Kits	Mouth mirror, dental probe, college tweezers	20	40	100	Adequate
Lead Screen		1	1	5	Adequate
Biopsy Kit		1	1	2	Adequate
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	1	2	1	Adequate
Computers	Minimum Pentium – IV	1	1	2	Adequate
	With High resolution	1	1	1	Adequate
Needle Burner with Syringe Cutter		2	2	2	Adequate

Department: Oral Pathology and Oral Microbiology

Name			Quantity		Adamina /
	Specification	50 Intake	100 Intake	Available Quantity	Adequate / Inadequate
Dental Chairs and Units	With shadowless lamp, spittoon, 3 way syringe, instrument tray and suction	1	2	3	Adequate
Microscopes		20	40	35	Adequate
Microtome		1	1	2	Adequate
Wax bath		1	1	2	Adequate
Water bath		1	1	2	Adequate
Knife sharpener		1	1	1	Adequate
Hot plate		1	1	2	Adequate
Spencer Knife		1	1	2	Adequate

Department: Public Health Dentistry

	Specification	Qua	ntity	Available	Adequate / Inadequate
Name		50 Intake	100 Intake	Quantity	
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventury suction, micromotor control light cure 3 way syringe, X-ray viewer, instrument tray Dental Operators stool with height adjustment with shadowless lamp, spittoon, 3 way syringe, instrument tray and suction, micromotor, airotor, light cure		16	11	Adequate
Autoclaves	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	1	2	2	Adequate
Ultrasonic cleaner	Minimum capacity 13 litres with mesh bucket with digital timer	1	2	2	Adequate
Needle burner with syringe cutter		2	4	2	Adequate
MOBILE CLINIC					
Mobile dental van	Mobile dental van with two dental chairs with all the attachments and adequate sitting space for 15 to 20 people			1	Adequate
Dental chair with unit	Hydraulically operated with spittoon attachment, halogen light with 2 intensity, air ventury suction, airrotor, micromotor, 3 way-scaler and light cure, x-ray viewer, instrument tray, operating stool	2		1	Adequate
Autoclaves	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	1	1	1	Adequate
Intraoral x-ray	Portable, 70 KV, 8 mA	1	1	1	Adequate
Glass bead sterilisers		1	1	1	Adequate
Compressor	1.25 HP	1	1	1	Adequate
Metal Cabinet	With wash basin	1	1	1	Adequate
Portable dental chair	Suitcase unit with airotor, micromotor, scaler and compressor 0.25 HP	2	2	2	Adequate
Stabiliser	4 KV	1	1	1	Adequate
Generator	4 KV	1	1	1	Adequate
Water tank	400 ltrs	1	1	1	Adequate
Oxygen cylinder		1	1	1	Adequate
Public address system		1	1	1	Adequate

TV and video cassette player	1	1	1	Adequate
Demonstration models	1	1	1	Adequate

Department: Paediatric and Preventive Dentistry

Name	Specification		ntity	Available	Adequate /
3000000000		50 Intake	100 Intake	Quantity	Inadequate
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power motorized evacuation system, Air rotor with miniature, Airrotor HPS, Micro motor with miniature contrangle Hand piece, 3 way syringe, ultrasonic sealer with 3 tips, Light cure unit LED based heat free, X-ray viewer, instrument tray Dental Operator'sstool with height adjustment (Pedo chair preferred)		20	10	Adequate
Autoclaves	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	1	2	1	Adequate
Ultrasonic cleaner	Minimum capacity 13 liters with mesh bucket with digital timer	1	2	1	Adequate
Needle Burner with syringe cutter		2	3	2	Adequate
Amalgamator		1	1	1	Adequate
Pulp Tester Digital		1	1	1	Adequate
Rubber dam kit for pedo		3	5	3	Adequate
Apex locator		1	1	1	Adequate
Endo motor	With torque control HPs	1	1	1	Adequate
Injectable gutta percha with condensation		1	1	1	Adequate
Radiovisiography	Digital intra X-ray system with pedo sensor and software	1	1	1	Adequate
Intra Oral Camera	With high resolution	1	1	- 1	Adequate
Scale instruments		5	10	5	Adequate
Restorative instruments		5	10	5	Adequate
Extraction forceps		5	10	5	Adequate
Intra-oral X-ray		1	1	1	Adequate
Automatic Developer		1	1	1	Adequate
Computer	Minimum Pentium – IV	1	1	2	Adequate
	PEDO LAB				1,000
Plaster dispenser	One each for plaster and stone plaster	2	2	2	Adequate
Model Trimmer	With diamond disc	1	1	1	Adequate
Model Trimmer	Double disc one diamond and one carborandam disc	1	1	1	Adequate
Welder with soldering attachments		1	1	1	Adequate
Vibrator		2	2	2	Adequate
Lab micro motor	Heavy duty with Hps	2	3	2	Adequate
Dental Lathe		1	1	1	Adequate
Model Trimmer		1	1	1	Adequate

Steam cleaner	1	1	1	Adequate
Pressure moulding machine	1	1	1	Adequate
Carborandum disc	1	1	1	Adequate
Diamond disc	1	1	1	Adequate

Department: Orthodontics

		Qua	ntity	Acceptance	A.1.
Name	Specification	50 Intake	100 Intake	Available Quantity	Adequate / Inadequate
Dental Chairs and Units	Light with 2 intensity, Air Ventury suction, micromotor control light cure 3 way syringe, X-ray viewer, instrument tray Dental Operators stool with height adjustment	9	18	17	Adequate
'Init mount scaler		3	5	3	Adequate
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	1	2	2	Adequate
Ultrasonic cleaner	Minimum capacity 13 litres with mesh bucket with digital timer	1	2	1	Adequate
ORTHO LAB					Adequate
Plaster dispenser	One each for plaster and stone plaster	2	2	2	Adequate
Vibrator		2	2	2	Adequate
Model Trimmer		1	2	2	Adequate
Micromotor	Heavy duey	2	4	3	Adequate
Lathe		1	2	2	Adequate
X-ray viewers		2	2	3	Adequate
cephalostat	If available in radiology its is adequate	1	1	1	Adequate
Welders		2	4	3	Adequate
Blue Torch		1	1	. 1	Adequate
Base Formers		2	4	3	Adequate
podont		2	4	3	Adequate
Set of pliers		5	10	5	Adequate
Welder with soldering attachments		1	1	1	Adequate
Hydro solder		1	1	1	Adequate
Typhodont articulator	With metal teeth wax rim of Class I, II, III	3	4	3	Adequate
Pressure moulding machine		1	1	1	Adequate

Department: Periodontology

		Qua	ntity		Ad
Name	Specification	50 Intake	100 Intake	Available Quantity	Adequate / Inadequate
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evacuation system, Air Ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contrangle Handpiece, with shadowless lamp, spittoon, 3 way syringe, Ultrasonic sealer with 3 tips, X-ray viewer, instrument tray Dental Operators stool with height adjustment		34	21	Adequate
Scaling instrument sets		5	10	10	Adequate
Surgical instrument sets		3	6	6	Adequate
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	2	3	4	Adequate
Ultrasonic scaler	Minimum capacity 13 litres with mesh bucket	1	2	2	Adequate
lectro surgical cautery		1	1	1	Adequate
Needle burner with syringe cutter		4	6	4	Adequate
LASER	Soft tissue laser	1	1	2	Adequate
Surgical motor with physio dispenser		1	1	1	Adequate

Department: Conservative Dentistry and Endodontics

		Quantity			ENGLOS PROGRAMMENT
Name		50 Intake	100 Intake	Available Quantity	Adequate / Inadequate
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evacuation system, Air Ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contrangle Handpiece, with shadowless lamp, spittoon, 3 way syringe, Ultrasonic sealer with 3 tips, X-ray viewer, instrument tray Dental Operators stool with height adjustment		34	25	Adequate
Rubber dam kits		4	6	4	Adequate
Restorative instruments kits		5	10	6	Adequate
R.C.T. instrument kits		5	10	5	Adequate
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	2	3	2	Adequate
Ultrasonic cleaner	Minimum capacity 13 litres with mesh bucket	1	2	1	Adequate
Needle Burner with Syringe Cutter		3	4	3	Adequate
Amalgamator	With auto proportion, auto dispenser	2	3	2	Adequate
Rubber dam kits		4	6	4	Adequate
Pulp Tester Digital		2	4	2	Adequate
Apex locator		1	2	1	Adequate
Glass bead		4	6	4	Adequate

Sterilizers					
Plaster dispenser		2	2	2	Adequate
Vibrator		1	2	1	Adequate
Ceramic Unit		1	1	1	Adequate
Casting machine		1	1	1	Adequate
Intra-oral X-ray Unit	Proper radiation safety	1	1	1	Adequate
Automatic Developer		1	1	1	Adequate
Radiovisiography	RVG with Computer	1	1	1	Adequate
Endo motor	With torque control Hps	1	1	1	Adequate
Bleaching Unit		1	1	1	Adequate
Magnification loops		1	2	1	Adequate
Injectable gutta percha		1	2	1	Adequate
PHANTOM LAB UNIT	Phantom Table fitted with Halogen Operating Light Phantom Head body type neck joint for all the movement, TMJ movement, Modular with Air rotor, Micro motor with contra angle Hps, 3 way syringe, jaw with ivorine teeth, preferably soft gingival, dental operator's stool (not to use extracted or cadaver teeth).	30	60	30	Adequate

Chemical Laboratory

	and the second s	Quantity		A !! - ! - ! - !	0 da
Name	Specification	50 Intake	100 Intake	Available Quantity	Adequate / Inadequate
Plaster Dispenser	One each for plaster and stone plaster	2	2	2	Adequate
Model Trimmer	Carborandum disc	1	1	1	Adequate
	Diamond disc	1	1	1	Adequate
Lathe	Heavy Duty	1	2	1	Adequate
Lab Micromotor	With heavy duty handpiece	2	3	2	Adequate
Ultrasonic cleaner	Minimum capacity 5 litres	1	1	1	Adequate
pindle Grinder		1	1 (1	Adequate
Vibrator		1	2	1	Adequate
Burnout furnace		1	1	1	Adequate
Porcelain furnace		1	1	1	Adequate
Sandblasting Machine		1	1	1	Adequate
Lab Airotor		1	1	1	Adequate
Pindex System		1	1	1	Adequate
Circular saw		1	1	1	Adequate
Vacuum mixer		1	1	1	Adequate
Pnematic chisel		1	1	1	Adequate
Casting machine	Motor cast / induction casting preferred	1	1	1	Adequate

Department: Oral & Maxillofacial Surgery

A) Exodontia

Name	Specification	Qua	ntity		Adequate / Inadequate
		50 Intake	100 Intake	Available Quantity	
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventury suction, X-ray viewer, 3 way syringe, instrument tray Dental Operators stool and height adjustment and suction, Micromotor / Airotor		30	15	Adequate
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	2	2	06	Adequate
Ultrasonic cleaner	Minimum capacity 13 litres with mesh bucket	1	2	02	Adequate
Needle Burner with Syringe Cutter		4	6	07	Adequate
Extraction forceps sets	Complete set	10	20_	20	Adequate
)ental elevators	Complete set	5	10	11	Adequate
Minor Oral Surgery kits		3	6	03	Adequate
Emergency drugs tray		1	1	03	Adequate
Oxygen cylinder with mask		1	1	03	Adequate
X-ray viewers		2	2	04	Adequate
Computer	Minimum Pentium – IV	1	1	02	Adequate

B) Minor Surgery

			ntity	A D - LT-	4.7
Name	Specification	Intake Intake Quantity	Available Quantity	The second secon	
Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evaculation system, Air ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contra-angle Handpiece, 3 way syringe, instrument tray Dental Operators stool and height adjustment and suction.		5	06	Adequate

Department: Prosthodontics and Crown & Bridge

			ntity		Adequate / Inadequate
Name Specification	50 Intake	100 Intake	Available Quantity		
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evacuation system, Air Ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contrangle Handpiece, 3 way syringe, Ultrasonic sealer with 3 tips, Light cure unit, instrument tray and suction, Dental Operators stool with height adjustment		34	17	Adequate
Semi-adjustable articulator	With face bow	2	2	2	Adequate
Extra oral / intra oral tracer		2	2	2	Adequate
Dewaxing unit		1	2	1	Adequate

Curing unit		1	2	1	Adequate
Dental casting machine		1	1	1	Adequate
Wax burnout furnace		1	1	1	Adequate
Pre heating furnace		1	1	1	Adequate
Surveying unit		1	2	1	Adequate
Heavy duty hand piece	Lab micromotors	3	4	3	Adequate
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	1	2	1	Adequate
Needle Burner with Syringe Cutter		1	2	1	Adequate
Plaster dispenser	One each for plaster and stone plaster	2	2	2	Adequate
Model Trimmer with Carborandum		1	1	1	Adequate
disc Model Trimmer with diamond disc		1	2	1	Adequate
Acryliser		2	3	2	Adequate
Lathe		1	2	1	Adequate
Flask press		4	4	4	Adequate
Deflasking unit		4	4	4	Adequate
Dewaxing unit		2	3	2	Adequate
Hydraulic Press		2	3	2	Adequate
Mechanical Press		1	2	1	Adequate
Vacuum mixing machine		1	1	1	Adequate
Lab Micro motor	With heavy duty handpiece	3	4	3	Adequate
Curing pressure		1	1	1	Adequate
orcelain furnace		1	2	1	Adequate
Vibrator		1	2	1	Adequate
Sandblasting unit		2	2	2	Adequate
Ultrasonic cleaner		1	2	1	Adequate
Model Trimmer		2	4	2	Adequate
Hot water sterilizer		1	2	1	Adequate
Geyser	Compound bath	1	2	1	Adequate
H.P. grinder with suction		2	3	2	Adequate
Heavy duty lathe		2	2	2	Adequate
Phantom heads		50	50	50	Adequate
Pre-clinical working tables	Gas connection & Bunsen burner	50	100	50	Adequate

Ceramic and Cast Partial Laboratory

A12/25		2000	ntity	Available	Adequate /
Name	Specification	50 Intake	100 Intake	Quantity	Inadequate
Plaster dispenser	One each for plaster and stone plaster	2	2	2	Adequate
Duplicator		1	1	1	Adequate
Pindex System		1	1	1	Adequate
Circular saw		1	1	1	Adequate
Burnout furnace		1	1	1	Adequate
Sandblasting Machine	With two containers	1	1	1	Adequate
Electro polisher		1	1	1	Adequate
Model Trimmer with Carborandum disc		1	1	1	Adequate
Model Trimmer with diamond disc		1	1	1	Adequate
Induction casting machine		1	1	1	Adequate
Programmable porcelain furnace with vacuum pump with instrument kit and material kit		1	1	1	Adequate
opot welder with soldering attachment of cable		1	1	1	Adequate
Vacuum mixing machine		1	1	1	Adequate
Steam Cleaner		1	1	1	Adequate
Spindle Grinder 24000 rpm with vacuum suction		1	1	1	Adequate
Wax heater		1	1	1	Adequate
Wax carver		1	1	1	Adequate
Curing pressure pot		1	1	1	Adequate
Milling machine		1	1	1	Adequate
Heavy duty lathe with suction		1	1	1	Adequate
Preheating furnace		1	1	1	Adequate
Palatal trimmer		1	1	1	Adequate
Ultrasonic cleaner	5 liters capacity	1	1	1	Adequate
Composite curing unit		1	1	1	Adequate
Micro surveyor		1	1	1	Adequate
re-clinical Prosthetic Laboratory	Work table preferably complete stainless steel fitted with light, Bunsen burner, air blower, working stool. Adequate number of lab micro motor with attached hand piece.	30	60	30	Adequate
Plaster room for pre-clinical work	• Visite 17				Adequate
Plaster dispenser	One each for plaster and stone	2	2	2	Adequate
Vibrator	proces and stone	2	2	2	Adequate
athe		2	2	2	Adequate
Model Trimmer		1	1	1	
Carborandum Disc		1	1	1	Adequate
Diamond disc		1	1	1	Adequate Adequate

EQUIPMENT REQUIREMENTS

As per MDS Course Regulations, 2017

Department: Prosthodontics and Crown & Bridge

Name	lame Specification		tity	Available Quantity	Adequate /
Electrical Dental Chairs and Units	motorized suction, micromotor and airotor	unit per PG student and two chairs with		20	Adequate
		1 Unit		12	Adequate
Articulators – semi adjustable/ adjustable with face bow		6	12	- 6	Adequate
Micromotor – (Lab Type can also be attached 'fixed) to wall		2	6	2	Adequate
Ultrasonic scaler		2	2	2	Adequate
Light cures		2	2	1	Adequate
Hot air oven		1	1	2	Adequate
Autoclave		2	2	2	Adequate
Surveyor		2	2	1	Adequate
Refrigerator		1	1	2	Adequate
X-ray viewer		1	2	3	Adequate
Pneumatic, Crown bridge remover		2	3	2	Adequate
Needle destroyer		1	2	1	Adequate
Intra oral camera		1	1	1	Adequate
Digital SLR camera		1	1	1	Adequate
Computer with internet connection with attached printer and scanner		1	1	1	Adequate
LCD projector		1	1	1	Adequate
Clinical Lab for Prosth	etics			-2	
Plaster dispenser		2	2	2	Adequate
odel trimmer with carborandum Disc		1	2	2	Adequate
Model trimmer with diamond disc		1	2	3	Adequate
High speed lathe		2	3	4	Adequate
Vibrator		2	4	2	Adequate
Acrylizer		1	2	2	Adequate
Dewaxing unit		1	2	2	Adequate
Hydraulic press		1	2	1	Adequate
Mechanical press		1	1	1	Adequate
Vacuum mixing machine		1	1	3	Adequate
Micro motor lab type		2	3	1	Adequate
Curing pressure pot		1	1	1	Adequate
Pressure molding machine		1	1	1	Adequate
Chrome - Cobalt Lab Equi	pment	S	1- 1		
Duplicator		1	1	1	Adequate
Pindex system		1	1	1	Adequate
Burn-out furnace		1	1	1	Adequate
Welder		1	1	1	Adequate
	Micro and macro	1	1	1	Adequate
Electro – polisher		1	1	1	Adequate
Model trimmer with carborandum disc		1	1	12	Adequate

Model trimmer with diamond disc		1	1	1	Adequate
Model trimmer with double disc (one		1	1	1	Adequate
Carborandum and one	tod the state				PAUL BOOK STATES
diamond disc)					
Casting machine, motor cast with		1	1	1	Adequate
the safety door closure, gas blow torch with					
regulator					
Dewaxing furnace		1	1	1	Adequate
Induction casting machine with vacuum		1	1	1	Adequate
pump, capable of casting					
titanium chrome cobalt precision metal					
Spot welder with soldering, attachment of		1	1	1	Adequate
cable					
Steam cleaner		1	1	1	Adequate
Vacuum mixing machine		1	1	1	Adequate
Spindle grinder 24,000 ROM with vacuum		1	1	1	Adequate
suction				4/2	
Wax heater		2	3	3	Adequate
Vax carvers (Full PKT Set)		2	3	3	Adequate
Milling machine		1	1	2	Adequate
Stereo microscope		1	1	1	Adequate
Magnifying work lamp		1	1	1	Adequate
Heavy duty lathe with suction		1	1	1	Adequate
Preheating furnace		1	1	1	Adequate
Dry model trimmer		1	1	1	Adequate
Die cutting machine		1	2	2	Adequate
Ultrasonic cleaner		1	1	1	Adequate
Composite curing unit		1	1	1	Adequate
Ceramic Lab Equipm	ent				
Fully programmable porcelain furnace with		1	1	2	Adequate
vacuum pump					
Ceramic kit (instruments)		3	3	3	Adequate
Ceramic materials (kit)		1	1	1	Adequate
Ceramic polishing kit		2	2	_2	Adequate
/ Implant Equipmen	t			18	
Electrical dental chair and unit		1	1	*1	Adequate
Physio dispenser		1	1	1	Adequate
Implant kit	Minimum 2 systems	2	2	2	Adequate
Implants		10	10	10	Adequate
Prosthetic components		10	10	10	Adequate
Unit mount light cure		1	2	2	Adequate
X-ray viewer		1	2	2	Adequate
Needle destroyer		1	2	2	Adequate
Ultrasonic cleaner capacity 3.5 lts		1	1	1	Adequate
Autoclave programmable for all recommended cycles		1	2	2	Adequate
X-ray machine with RVG		1	1	1	Adequate
Refrigerator		1	1	2	Adequate
Surgical kit/prosthetic kit		2	2	2	Adequate
Educating models		1	1	1	Adequate
Implant removing instruments		1	1	1	Adequate

Department: Periodontology

Name	Specification		Quantity		Adequate / Inadequate	
Dental Chairs and Units	Electrically operated with shadowless lamp, spittoon, 3 way syringe, instrument tray and motorized suction, micromotor attachment with contra angle handpiece, airoter attachment, ultrasonic scaler (Piezo) with detachable autoclavable hand piece	mp, unit per post- graduate student and Two chairs ent with unit for the faculty onic vith ble		n, unit per post- graduate student and Two chairs with unit for the faculty		
\uto clave (fully automatic) front loading		1 Unit	2 Units			
Steel bin		1	2	2	Adequate	
Airoter hand pieces		4	6	6	Adequate	
UV chamber		2	2	4	Adequate	
Formalin chamber		1	1	2	Adequate	
W.H.O probe		1	1	1	Adequate	
Nabers probe		2	2	2	Adequate	
Williams probe		2	2	2	Adequate	
UNC-15 probe		2	2	2	Adequate	
Gold Man fox probe	-	4	4	4	Adequate	
Pressure sensitive probe		1	1	1	Adequate	
Marquis color coded probe		1	1	1	Adequate	
Supra gingival scalers		1	1	1	Adequate	
Sub gingival scaler	set	2	2	4	Adequate	
Arkansas sharpening stone	set	2	2	4	Adequate	
		1	1	1	Adequate	
Surgical Instrume Coutine surgical instrument kit (Benquis						
reriosteal elevator, periotome)	set	2	3	2	Adequate	
Surgery trolleys		6	6	6	Adequate	
X ray viewer		1	2	3	Adequate	
Surgical cassette with sterilisation pouches Electro surgery unit		4	6	4	Adequate	
		1	1	1	Adequate	
Special Surgical Instru Kirkland's knife						
Orban's knife	set	1	1	10	Adequate	
Paquette blade handle	set	1	1	10	Adequate	
Krane kaplan pocket marker		1	1	1	Adequate	
Mc Calls universal curettes	set	1	1	10	Adequate	
Gracey's curettes (No.1-18)	set	1	1	1	Adequate	
Mini five curettes	set	2	2	12	Adequate	
Cumine scalar	set	1	1	1	Adequate	
Mallet		1	1	20	Adequate	
Chisel		1	1	2	Adequate	
Oschenbein chisel	straight averal	1	1	2	Adequate	
Schluger bone file	straight, curved	1	1	1	Adequate	
Bone fixation screw kit		1	1	1	Adequate	
Bone scrapper		1	1	1	Adequate	
The scrapper		1	1	1	Adequate	

Bone trephines for harvesting autografts	1 set	1	1	1	Adequate
Bone regenerative materials	Bone graft and GTF membranes	5	5	5	Adequate
Local drug delivery systems	At least two different agents	1005	1	1	Adequate
Root conditioning agent	At least two different agents	2	2	2	Adequate
Micro needle holder	1177 (2400-04)	1	1	1	Adequate
Micro scissors		1	1	1	Adequate
Magnifying loop (2.5 – 3.5)		1	2	1	Adequate
Operating microscope	optional	1	1		Adequate
3rd generation digital probe	optional	1	1	+	Adequate
Bone expander and bone crester	optional	1	1	1	Adequate
Distraction osteogenesis kit	optional	1	1		Adequate
Bone mill	optional	1	1	1	Adequate
Bone graft / membrane placement spoon		1	1	1	
Bone condenser		1	1.	1	Adequate
Peizo-surgery unit	optional	1	1	1	Adequate
entrifuge for PRP/PRF preparation	optional	1	1	1	Adequate
Soft tissue laser (8 watt)	operation.	1	control .	1	Adequate
Osteotome	set optional	1	1	3	Adequate
Miscellaneous Instrum		1	1	1	Adequate
Composite gun with material kit	il Citto	1			Adequate
Splinting kit with material		1	1	1	Adequate
Composite finishing kit		2	3	2	Adequate
Glass Ionomer cement		1	1	1	Adequate
Digital camera		1	1	1	Adequate
Intra Oral camera		1	1	1	Adequate
Ultrasonic cleaner		1	1	1	Adequate
Emergency kit		1	1	1	Adequate
Refrigerator		1	1	1	Adequate
X-ray viewer		1	1	2	Adequate
LCD projector		2	2	2	Adequate
TO THE PARTY OF TH		1	1	1	Adequate
Computer with internet nnection with attached printer and scanner		1	1	1	Adequate
mplant Equipment				34	Adequate
lectrical dental chair and unit		1	1	1	Adequate
Physio dispenser		1	1	1	Adequate
mplant kit	At least two different systems	2	2	2	Adequate
mplants		10	10	10	Adequate
mplant maintenance kit (plastic instruments)		1 set	1 set	1	Adequate
mplant guide		1	1		Adequate
(-ray viewer		1	2	1	Adequate
Needle destroyer		1	2	1	Adequate
Ultrasonic cleaner capacity 3.5 lts		1	1	1	Adequate
autoclave programmable for all ecommended cycles		1	1	1	Adequate
VG with x-ray machine		1	1	1	Adequate
efrigerator		1	1	1	Adequate
urgical kit		2	2	1	Adequate
inus lift kit		1	1	1	Adequate
ducating models		1	1	1	Adequate
mplant removing kit		1	1	1	Adequate

Department: Oral & Maxillofacial Surgery

Name	Specification	Quantity		Quantity		Available Quantity	Adequate Inadequate
Dental Chairs and Units	Electrically operated with one chair and shadowless lamp, spittoon, 3 way syringe, graduate instrument tray and high student and						
	otorized suction, with micromotor and	Two with	chairs unit for				
	micro motor attachment	the fac	ulty				
Autoclave			2 Units				
	Front loading	2	3	8	Adequate		
Fumigators	N	1	1	1	Adequate		
Oscillating saw	With all hand pieces	1	1	1	Adequate		
Surgical Instrument	S						
General surgery kit including tracheotomy kit Minor oral surgery kit		2	2	2	Adequate		
Osteotomy kit		5	10	5	Adequate		
Cleft surgery kit		1	1	1	Adequate		
Bone grafting kit		1	1	1	Adequate		
		1	1	1	Adequate		
Emergency kit		1	1	1	Adequate		
Trauma set including bone		2	2	2	Adequate		
Plating kit	12 1020 O	1	1	1	Adequate		
Implantology kit	Minimum 2	2	2	2	Adequate		
With implants	systems	10	10	48	Adequate		
Distraction osteogenesis kit		1	1	1	Adequate		
Peizo surgical unit		1	1	1	Adequate		
Magnifying loops		1	1	1	Adequate		
Operating microscope and Microsurgery kit	desirable	1	1	1	Adequate		
Dermatomes		1	1	1	Adequate		
Needle destroyer		2	3	4	Adequate		
Ultrasonic Cleaner capacity 3.5 lts		1	1	2	Adequate		
ormalin chamber		1	1	. 1	Adequate		
Pulse oxymeter		1	1	4	Adequate		
Ventilator		1	1	- 1	Adequate		
Major operation theatre with all facilities		1	1	1	Adequate		
Recovery and Intensive Care Unit with all necessary life support equipments		2 beds	2 beds	2 beds	Adequate		
Fibrooptic light		1	1	1	Adequate		
npatient beds		20	20	20	Adequate		
Fiber optic laryngoscope		1	1	1	Adequate		
Computer with internet connection with attached printer and scanner		1	1	2	Adequate		
-CD projector		1	1	1	Adequate		
Refrigerator		1	1	2	Adequate		
					Auequate		

Department: Conservative Dentistry and Endodontics

Name	Specification	unit per post- graduate student and two chairs with		Available Quantity	Adequate / Inadequate	
Dental Chairs and Units	shadowless lamp, spittoon, 3 way syringe, instrument tray and motorized suction,			adowless lamp, unit per post- troon, 3 way syringe, graduate student and two chairs with cromotor, airotor achment with	shadowless lamp, spittoon, 3 way syringe, instrument tray and motorized suction, micromotor, airotor attachment with unit for the faculty	er post- ite it and airs with for the
ENDOCONICHANDOLEGE		1 Unit				
ENDOSONIC HANDPIECES – Micro endosonic Tips, retro treatment		2	2	2	Adequate	
Mechanised rotary instruments including hand pieces (speed and torque control) and hand instruments various systems		3	3	3	Adequate	
Rubber dam kit		1 per	9	9	Adequate	
Autoclaves for bulk instrument sterilization vacuum (Front loading)		2	2	2	Adequate	
Autoclaves for hand piece sterilization		1	1	1	Adequate	
Apex locators one for every two chairs		2	2	2	Adequate	
Pulp tester		2	2	2	Adequate	
Equipments for injectable thermoplasticized gutta percha		1	1	1	Adequate	
Operating microscopes 3 step or 5 step magnification		1	1	1	Adequate	
Surgical endo kits (Microsurgery)		2	2	2	Adequate	
Set of hand instruments (specifications required)		1	1	1	Adequate	
Sterilizer trays for autoclave		4	4	4	Adequate	
Ultrasonic cleaner capacity 3.5 lts		1	1	. 1	Adequate	
Variable Intensity polymerization equipments - VLC units	Desirable	1	1	1	Adequate	
Conventional VLC units one for every two chairs		2	2	2	Adequate	
Needle destroyer		2	2	2	Adequate	
Magnifying loupes one for students and one for faculty		1	2	2	Adequate	
LCD projector		1	1	1	Adequate	
Composite kits with different shades and polishing kits		2	2	2	Adequate	
Ceramic finishing kits, metal finishing kits	In ceramic labs	2	2	2	Adequate	
Amalgam finishing kits		2	2	2	Adequate	
RVG with x-ray machine developing kit		1	1	1	Adequate	
Chair side micro abrasion		1	1	1	Adequate	
Bleaching unit		1	1	1	Adequate	
Instrument retrieval kits		1	1	1	Adequate	
Computer with internet connection with attached printer and scanner		1	1	1	Adequate	

Refrigerator	1	1	1	Adequate
Equipments for casting procedures				Adequate
Equipments for ceramics including induction casting machines/ burnout preheat furnaces/ wax elimination furnaces	1	1	1	Adequate
Lab micro motor/ metal grinders / sand blasters/ polishing lathes/ duplicator equipment/ vacuum investment equipments	1	1	1	Adequate
Laser (preferably hard tissue)	1	1	1	Adequate
Face bow with semi adjustable articulator	1	2	1	Adequate

Department: Orthodontics and Dentofacial Orthopedics

Name	Specification Quar		Quantity		Adequate Inadequate
Dental Chairs and Units	Electrically operated with shadow less lamp, spittoon, 3 way syringe, instrument tray and motorized suction	unit studen Two	chair - & per PG t and chairs unit for		
		1 Unit	2 Units		
Vacuum /pressure moulding unit		1	1	1	Adequate
Hydrogen soldering unit		1	1	1	Adequate
Lab micromotor		3	5	3	Adequate
Spot welders		3	5	3	Adequate
Model trimmer (Double disc)		2	3	2	Adequate
Light curing unit		2	2	2	Adequate
High intensity light curing unit		1	2	2	Adequate
Polishing lathes		2	3	2	Adequate
Tracing tables		3	5	3	Adequate
SLR digital camera		1	1	. 1	Adequate
canner with transparency adapter		1	1	1	Adequate
X-ray viewer		3	4	3	Adequate
LCD projector		1	1	1	Adequate
Autoclaves for bulk instrument Sterilization vacuum (Front loading)		1	1	2	Adequate
Needle destroyer		1	1	1	Adequate
Dry heat sterilizer		1	1	2	Adequate
Ultrasonic scaler		1	1	3	Adequate
Sets of Orthodontic pliers		3	3	5	Adequate
Orthodontic impression trays		3	5	3	Adequate
Ultrasonic cleaner capacity 3.5 lts		1	1	1	Adequate
Electropolisher		1	1		Adequate
Typodonts with full teeth set		3	3		Adequate
Anatomical articulator with face bow attachments		1	1	1	Adequate
Free plane articulators		1	1	1	Adequate
Hinge articulators		4	4	100	Adequate
Computer software for cephalometrics		1	1		Adequate
Computer with internet connection with attached printer and scanner		1	1		Adequate
Refrigerator		1	1	2	Adequate

Department: Oral & Maxillofacial Pathology and Oral Microbiology

Name	Specification	Quantity		Available Quantity	Adequate /
		1 Unit	2 Units		
Dental Chairs and Units	Electrically operated with shadow less lamp, spittoon, 3 way syringe, instrument tray and suction		6	3	Adequate
required for processing of biopsy specimens and staining.	Reasonable quantity should be made available			24	Adequate
Adequate tissue capsules / tissue embedding cassettes	Reasonable quantity should be made available			24	Adequate
Paraffin wax bath	thermostatically controlled	1	1	2	Adequate
Leuckhart pieces		10	10	24	Adequate
Block holders		25	25	40	Adequate
1icrotome	Manual	1	1	01	Adequate
Microtome	semi – automated	1	1	01	Adequate
Tissue floatation water bath	thermostatically controlled	1	1	02	Adequate
Slide warming table		1	1	02	Adequate
Steel slide racks for staining		5	5	05	Adequate
Diamond glass marker		2	2	02	Adequate
Research microscope with phase contrast, dark field, polarization, image analyzer, photomicrography attachments		1	1	01	Adequate
Multi head microscope	Penta headed	1	1	01	A -1
Binocular compound microscope	7	2 for	4 for	11	Adequate
		faculty and one per student	faculty and one per student		Adequate
Stereo microscope		1	1	01	Adequate
^ luminum slide trays		5	5	.05	Adequate
wooden / plastic slide boxes		5	5	05	Adequate
Wax block storing cabinet		5,000 capacity	10,000 capacity	01	Adequate
Slide storing cabinet		5,000	10,000 capacity	02	Adequate
Refrigerator		1	1	01	Adequate
Pipettes		5	5	12	Adequate
Surgical kit for biopsy		3	6	03	Adequate
mmuno histo chemistry lab		1	1	01	Adequate
Computer with Internet Connection with ittached orinter and scanner		1	1	01	Adequate
CD projector		1	1	02	Adequate
Desirable Equipme	ent				quate
ryostat		1	1		Adequate
luorescent microscope		1	1	01	Adequate
lard tissue microtome		1	1	-	Adequate
issue storing cabinet (frozen)		1	1	-	Adequate
Microwave		1	1	3	Adequate

Department: Public Health Dentistry

Name	Specification	unit per post- graduate student and one chair with unit for the faculty		Available Quantity	Adequate / Inadequate
Dental Chairs	spittoon, 3 way syringe, instrument tray and motorized suction, micromotor attachment				
		1 Unit	2 Units		
Extraction forceps		4 sets	6 sets	NA	NA
Filling instruments		4 sets	6 sets	NA	NA
Scaling instruments	Supra gingival scaling	4 sets	6 sets	NA	NA
Amalgamator		1	1	NA	NA
Pulp tester		1	1	NA	NA.
Autoclave		1	1	NA	NA
X-ray viewer		1	1	NA	NA
Instrument cabinet		1	1	NA	NA
LCD or DLP multimedia projector		1	1	NA	NA
Computer with internet connection with attached printer and scanner		1	1	NA	NA
For peripheral dental care or fiel	d programme				NA
Staff bus	u programme	1	1		NA
Mobile dental clinic fitted with at least 2 dental chairs with complete dental unit with fire extinguisher		1	1	NA	NA
Ultrasonic scaler,		1	2	·-NA	NA
Ultrasonic cleaner capacity 3.5 lts		1	1	NA	NA
Compressor	One with chair			NA	NA
Generator		1	1	NA	NA
Public address system, audio-visual aids		1	1	NA	NA
Television		1	1	NA	NA
Digital Versatile Disc Player		1	1	NA	NA
Instrument cabinet, emergency medicine kits, Blood pressure apparatus		1	1	NA	NA
Portable oxygen cylinder		1	1	NA	NA
Portable chair		1	1	NA	NA
Refrigerator		1	1	NA	NA *

Department: Paediatric and Preventive Dentistry

Name	Specification	n, unit per post- graduate n, student and Two chairs with unit for the faculty e		Available Quantity	Adequate Inadequate
Dental Chairs and Units	The state of the s				madequate
Dode extenti 6		1 Unit	2 Units		
Pedo extraction forceps sets		3	4	NA	NA
Autoclaves for bulk instrument sterilization vacuum (Front loading)		1	2	NA	NA
RVG with intra oral x-ray unit		1	1	NA	NA
Automatic developer		1	1	NA NA	NA NA
Pulp tester		2	3	NA	NA NA
Apex locator		1	1	NA	NA
Rubber dam kit	One set per student	1	1	NA	NA
njectable GP condenser		1	1	NA	NA NA
ndodontic pressure syringe		1	1	NA	NA NA
Glass bead steriliser		2	4	NA	NA
Spot welder		2	3	NA	NA.
Jltrasonic scalers		2	4	NA	NA NA
Needle destroyer		1	1	NA	NA
ormalin chamber		1	1	NA	NA NA
Iltrasonic cleaner capacity 3.5 lts		1	1	NA	NA NA
-ray viewer		2	3	NA	NA
malgamator		1	2	NA NA	NA NA
faster dispenser		2	2	-NA	
ental lathe		1	2	NA	NA NA
librator		2	3	NA NA	NA
ypodonts	One set per student	1	1	NA NA	NA NA
oldering unit		1	1	NA NA	NA NA
and pinching beak pliers		2 Sets	2 Sets	NA NA	NA NA
roximal contouring pliers		2	3	NA NA	NA NA
rown crimping pliers		2	3	NA NA	
ouble beak pliers anterior and posterior		2	3	NA NA	NA NA
ab micro motor		2	3	NA NA	NA NA
cryliser		1	2	NA NA	NA NA
lagnifying loupes		1	1	NA NA	NA NA
onscious sedation unit	Desirable	1	1	NA NA	NA NA
ulse oxymeter		1	1	NA NA	NA NA
nantom head table with tached Light, Airotor and micro motor	One set per each P.G. Student	1	1	NA NA	NA NA
omputer with internet connection with tached printer and scanner		1	1	NA	NA
D projector		1	1	NA	NA
frigerator		1	1	NA NA	NA NA

Department: Oral Medicine and Radiology

Name	Specification	unit per post-		Available Quantity	Adequate / Inadequate
Dental Chairs and Units	Electrically operated with shadowless lamp, spittoon, 3 way syringe, instrument tray and suction			10	Adequate
		1 Unit	2 Units		Adequate
RVG with intra oral radiography machine (FDA Approved)	55-70 Kvp with digital compatibility	1	1	1	Adequate
Extra oral radiography machine	100 kvp	1	1	1	Adequate
Panoramic radiography (OPG) machine with cephalometric and TMJ attachment with printer	Digital compatibility	1	1	1	Adequate
Intra-oral camera		1	2	1	Adequate
Pulp tester		2	4	2	Adequate
Autoclave		1	1	1	Adequate
Punch biopsy tool		2	3	2	Adequate
Biopsy equipment		1	2	2	Adequate
Surgical trolley		2	2	3	Adequate
Emergency medicines kit		1	1	1	Adequate
Extra oral cassettes with intensifying screens (Conventional and rare earth)		4	6	10	Adequate
Lead screens		2	2	5	Adequate
Lead aprons		2	2	4	Adequate
Lead gloves		2	2	2	Adequate
Radiographic filters (Conventional and rare earth)		1	1	1	Adequate
Dark room with safe light facility		1	1	1	Adequate
utomatic radiographic film processors		2	2	- 2	Adequate
Radiographic film storage lead containers		1	1	1	Adequate
Thyroid collars		1	1	1	Adequate
Digital sphygmomanometer		1	1	1	Adequate
Digital blood glucose tester		1	1	1	Adequate
Digital camera		1	1	1	Adequate
X-ray viewer boxes		2	3	5	Adequate
Lacrimal probes		2 sets	2 Sets	2 Sets	Adequate
Sialography cannula		2 sets	2 Sets	2 Sets	Adequate
Illuminated mouth mirror and probe		2	2	2	Adequate
Computer with internet connection with attached printer and scanner		1	1	1	Adequate
LCD projector		1	1	1	Adequate
Refrigerator		1	1	1	Adequate